

Analyte	Acronym	Chemical Abstract Services Registry Number (CASRN)	Minimum Detection Limit (MDL)(ppt)*	Minimum Reporting Level (MRL)(ppt) **
Hexafluoropropylene oxide dimer acid	HFPO-DA	13252-13-6	0.29	1.00
N-ethyl perfluorooctanesulfonamidoacetic acid	NEtFOSAA	2991-50-6	0.6	2.50
N-methyl perfluorooctanesulfonamidoacetic acid	NMeFOSAA	2355-31-9	0.46	3.00
Perfluorobutanesulfonic acid	PFBS	375-73-5	0.16	1.00
Perfluorodecanoic acid	PFDA	335-76-2	0.31	1.00
Perfluorododecanoic acid	PFDoA	307-55-1	0.23	2.00
Perfluoroheptanoic acid	PFHpA	375-85-9	0.29	2.00
Perfluorohexanesulfonic acid	PFHxS	355-46-4	0.39	1.00
Perfluorohexanoic acid	PFHxA	307-24-4	0.24	1.00
Perfluorononanoic acid	PFNA	375-95-1	0.37	2.00
Perfluorooctanesulfonic acid *	PFOS	1763-23-1	0.43	2.00
Perfluorooctanoic acid *	PFOA	335-67-1	0.4	1.00
Perfluorotetradecanoic acid	PFTA	376-06-7	0.33	1.00
Perfluorotridecanoic acid	PFTrDA	72629-94-8	0.26	2.00
Perfluoroundecanoic acid	PFUnA	2058-94-8	0.37	1.00
11-chloroeicosafuoro-3-oxaundecane-1-sulfonic acid	11Cl-PF3OUdS	763051-92-9	0.44	2.00
9-chlorohexadecafluoro-3-oxanone-1-sulfonic acid	9Cl-PF3ONS	756426-58-1	0.38	2.00
4,8-dioxa-3H-perfluorononanoic acid	ADONA	919005-14-4	0.33	1.00
* Minimum Detection Limit (MDL) determined by the Maryland Department of Health Laboratories Administration- Division of Environmental Services				
** Minimum Reporting Level (MRL) determined by the Maryland Department of Health Laboratories Administration- Division of Environmental Services				
***Results that are reported as "ND" stands for "Non-Detect," meaning there may be levels of PFAS in the sample but at levels below the reporting limits.				